

Calibrating Unit New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00368	Calib. Date: 10/12/2012
Certification File No.:	00369	Cert. No.: 00018
Linearity File No.:	00371	Cert. Date: 10/12/2012
Solution File No.:	00389	Lin. Date: 10/12/2012
Sequential File No.:	00389	Lin. No.: 00014
		Soln. Date: 01/12/2013
		Soln. No.: 00084
		File Date: 01/12/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0378

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:17S		
Control 1 EC	0.100%	10:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		
Control 2 EC	0.098%	10:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		
Control 3 EC	0.099%	10:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE	First Name: BRYAN	MI: F
Signature: <u></u>	Badge No.: 96	Date: 01/12/2013

