

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00543	Calib. Date:	03/23/2015
Certification File No.:	00544	Cert. No.:	00017
Linearity File No.:	00545	Lin. Date:	03/23/2015
Solution File No.:	00548	Soln. No.:	00123
Sequential File No.:	00548	File Date:	04/18/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1239

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:39D		
Control 1 EC	0.100%	02:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	02:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40D		
Control 2 EC	0.101%	02:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	02:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:42D		
Control 3 EC	0.101%	02:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SNYDER

First Name: TIMOTHY

MI: J

Signature: \_\_\_\_\_



Badge No.: 105

Date: 04/18/2015