

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00445 Calib. Date: 09/19/2013 Calib. No.: 00020
Certification File No.: 00446 Cert. Date: 09/19/2013 Cert. No.: 00013
Linearity File No.: 00447 Lin. Date: 09/19/2013 Lin. No.: 00016
Solution File No.: 00472 Soln. Date: 02/08/2014 Soln. No.: 00106
Sequential File No.: 00472 File Date: 02/08/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0399

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:26S		
Control 1 EC	0.100%	17:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:27S		
Control 2 EC	0.097%	17:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:29S		
Control 3 EC	0.098%	17:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Sgt. [Signature]
#096

Badge No.: 96

Date: 02/08/2014

