

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00445 Calib. Date: 09/19/2013 Calib. No.: 00020  
Certification File No.: 00446 Cert. Date: 09/19/2013 Cert. No.: 00013  
Linearity File No.: 00447 Lin. Date: 09/19/2013 Lin. No.: 00016  
Solution File No.: 00469 Soln. Date: 01/18/2014 Soln. No.: 00105  
Sequential File No.: 00469 File Date: 01/18/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164  
Control Solution %: 0.100% Expires: 06/28/2014  
Solution Control Lot: 12F102 Bottle No.: 0398

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 21:50S |                |                     |
| Control 1 EC      | 0.099% | 21:51S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 21:51S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:52S |                |                     |
| Control 2 EC      | 0.097% | 21:53S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.098% | 21:53S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:53S |                |                     |
| Control 3 EC      | 0.097% | 21:54S | 34.1°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.098% | 21:54S | 34.1°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:55S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: SZYMANSKI

First Name: NICHOLAS

MI: J

Signature: \_\_\_\_\_

Badge No.: 086

Date: 01/18/2014

