

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00815	Calib. Date: 09/05/2019
Certification File No.:	00816	Calib. No.: 00042
Linearity File No.:	00817	Cert. Date: 09/05/2019
Solution File No.:	00848	Lin. No.: 00036
Sequential File No.:	00848	Soln. Date: 02/02/2020
		Soln. No.: 00218
		File Date: 02/02/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 1113

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:06S		
Control 1 EC	0.100%	17:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:08S		
Control 2 EC	0.101%	17:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	17:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:09S		
Control 3 EC	0.100%	17:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STATTEL

First Name: DAVID

MI: K

Signature: _____

PTM #116

Badge No.: 116

Date: 02/02/2020

