

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00965 Calib. Date: 12/20/2021 Calib. No.: 00048
Certification File No.: 00966 Cert. Date: 12/20/2021 Cert. No.: 00038
Linearity File No.: 00967 Lin. Date: 12/20/2021 Lin. No.: 00042
Solution File No.: 00983 Soln. Date: 02/25/2022 Soln. No.: 00263
Sequential File No.: 00983 File Date: 02/25/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 06/30/2022
Solution Control Lot: 20310 Bottle No.: 0484

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:02S		
Control 1 EC	0.099%	17:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:03S		
Control 2 EC	0.099%	17:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:05S		
Control 3 EC	0.099%	17:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PETERSON

First Name: BRIAN

MI: D

Signature: _____

Badge No.: 129

Date: 02/25/2022

