

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00864	Calib. Date: 05/04/2020
Certification File No.:	00865	Calib. No.: 00044
Linearity File No.:	00866	Cert. Date: 05/04/2020
Solution File No.:	00871	Cert. No.: 00038
Sequential File No.:	00871	Lin. Date: 05/04/2020
		Lin. No.: 00038
		Soln. Date: 06/03/2020
		Soln. No.: 00226
		File Date: 06/03/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	19150	Expires: 06/10/2021
		Bottle No.: 0945

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:54D		
Control 1 EC	0.098%	15:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	15:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:55D		
Control 2 EC	0.097%	15:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	15:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:56D		
Control 3 EC	0.096%	15:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 090

Date: 06/03/2020

