

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00864 Calib. Date: 05/04/2020 Calib. No.: 00044  
Certification File No.: 00865 Cert. Date: 05/04/2020 Cert. No.: 00034  
Linearity File No.: 00866 Lin. Date: 05/04/2020 Lin. No.: 00038  
Solution File No.: 00878 Soln. Date: 07/16/2020 Soln. No.: 00229  
Sequential File No.: 00878 File Date: 07/16/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164  
Control Solution %: 0.100% Expires: 06/10/2021  
Solution Control Lot: 19150 Bottle No.: 1049

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:45D		
Control 1 EC	0.097%	11:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	11:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		
Control 2 EC	0.096%	11:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		
Control 3 EC	0.096%	11:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	11:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:49D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: STACEY

First Name: BRANDON

MI: J

Signature:  # 78

Badge No.: 78

Date: 07/16/2020