

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00038
Calibration File No.:	00739	Cert. Date:	07/02/2018
Certification File No.:	00740	Lin. Date:	07/02/2018
Linearity File No.:	00741	Soln. Date:	07/23/2018
Solution File No.:	00747	File Date:	07/23/2018
Sequential File No.:	00747		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	0310

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:15D		
Control 1 EC	0.100%	09:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:17D		
Control 2 EC	0.099%	09:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18D		
Control 3 EC	0.099%	09:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 090

Date: 07/23/2018

