

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00927 Calib. Date: 03/09/2021 Calib. No.: 00046  
Certification File No.: 00928 Cert. Date: 03/09/2021 Cert. No.: 00036  
Linearity File No.: 00929 Lin. Date: 03/09/2021 Lin. No.: 00040  
Solution File No.: 00958 Soln. Date: 07/25/2021 Soln. No.: 00257  
Sequential File No.: 00958 File Date: 07/25/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 1120

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:19D		
Control 1 EC	0.098%	02:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:20D		
Control 2 EC	0.097%	02:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:22D		
Control 3 EC	0.097%	02:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	02:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETERSON

First Name: BRIAN

MI: D

Signature: \_\_\_\_\_

*[Handwritten Signature]* #129

Badge No.: 129

Date: 07/25/2021

