

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00864 Calib. Date: 05/04/2020 Calib. No.: 00044
Certification File No.: 00865 Cert. Date: 05/04/2020 Cert. No.: 00034
Linearity File No.: 00866 Lin. Date: 05/04/2020 Lin. No.: 00038
Solution File No.: 00888 Soln. Date: 09/15/2020 Soln. No.: 00233
Sequential File No.: 00888 File Date: 09/15/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 1064

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:14D		
Control 1 EC	0.097%	13:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15D		
Control 2 EC	0.096%	13:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	13:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:16D		
Control 3 EC	0.096%	13:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: _____

Badge No.: 090

Date: 09/15/2020

