

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00815	Calib. Date:	09/05/2019
Certification File No.:	00816	Cert. Date:	09/05/2019
Linearity File No.:	00817	Lin. Date:	09/05/2019
Solution File No.:	00821	Soln. Date:	09/25/2019
Sequential File No.:	00821	File Date:	09/25/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0795

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:10D		
Control 1 EC	0.099%	15:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12D		
Control 2 EC	0.099%	15:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13D		
Control 3 EC	0.100%	15:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: 

Badge No.: 090

Date: 09/25/2019