

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00036
Calibration File No.:	00681	Calib. Date:	07/27/2017
Certification File No.:	00682	Cert. Date:	07/27/2017
Linearity File No.:	00683	Lin. Date:	07/27/2017
Solution File No.:	00689	Soln. Date:	09/29/2017
Sequential File No.:	00689	File Date:	09/29/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	1053

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:02D		
Control 1 EC	0.101%	09:02D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:02D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03D		
Control 2 EC	0.100%	09:04D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:04D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04D		
Control 3 EC	0.100%	09:05D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:05D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LOYER

First Name: JOHN

MI: B

Signature: _____



Badge No.: 057

Date: 09/29/2017

