

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00633	Calib. Date: 09/22/2016
Certification File No.:	00634	Calib. No.: 00033
Linearity File No.:	00635	Cert. Date: 09/22/2016
Solution File No.:	00656	Cert. No.: 00024
Sequential File No.:	00656	Lin. Date: 09/22/2016
		Lin. No.: 00028
		Soln. Date: 01/23/2017
		Soln. No.: 00152
		File Date: 01/23/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	201605D	Expires: 05/17/2018
		Bottle No.: 0201

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:05S		
Control 1 EC	0.097%	23:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07S		
Control 2 EC	0.097%	23:07S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:07S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:08S		
Control 3 EC	0.096%	23:09S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	23:09S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BISE

First Name: N

MI: E

Signature: _____

Sgt. John Bise #60

Badge No.: 60

Date: 01/23/2017

JAN 23 2017