

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00965	Calib. Date:	12/20/2021
Certification File No.:	00966	Cert. Date:	12/20/2021
Linearity File No.:	00967	Lin. Date:	12/20/2021
Solution File No.:	00969	Soln. Date:	12/23/2021
Sequential File No.:	00969	File Date:	12/23/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	20310	Expires:	06/30/2022
		Bottle No.:	0367

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:48S		
Control 1 EC	0.099%	17:49S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:50S		
Control 2 EC	0.098%	17:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51S		
Control 3 EC	0.099%	17:52S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-066 (NM)

Changed By:

Last Name: MIMIKOS

First Name: NICHOLAS

MI: E.

Signature: TPRII. AA #7413

Badge No.: 7413

Date: 12/23/2021