

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00034
Calibration File No.:	00657	Calib. Date:	02/02/2017
Certification File No.:	00658	Cert. Date:	02/02/2017
Linearity File No.:	00659	Lin. Date:	02/02/2017
Solution File No.:	00661	Soln. Date:	02/23/2017
Sequential File No.:	00661	File Date:	02/23/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	0355

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:57S		
Control 1 EC	0.101%	09:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59S		
Control 2 EC	0.099%	09:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00S		
Control 3 EC	0.099%	10:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOYER

First Name: JOHN

MI: B

Signature: _____ 

Badge No.: 057

Date: 02/23/2017

