

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00481	Calib. Date:	04/10/2014
Certification File No.:	00482	Cert. No.:	00023
Linearity File No.:	00483	Cert. Date:	04/10/2014
Solution File No.:	00484	Lin. Date:	04/10/2014
Sequential File No.:	00484	Soln. Date:	04/10/2014
		File Date:	04/10/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1231

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:08D		
Control 1 EC	0.099%	17:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:10D		
Control 2 EC	0.099%	17:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:11D		
Control 3 EC	0.099%	17:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDVJ P2-066 (15)

**Changed By:**

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SGT. [Signature] #5925

Badge No.: 5925

Date: 04/10/2014