

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Marlboro Twp. Police Dept. 1979 Township Drive Marlboro, NJ 07746		2. Contact: Sgt. Bryan Ohare	4. Date: 4/10/2014
		3. Phone Number: (732) 536-0100	5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXA-0069</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Returned from Draeger.

10. Comments:
Instrument was put back into service and a Calibration was performed on same with positive results.


11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

12. The above Instrument/Component is placed out of service pending further evaluation.

13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Marc Dennis #5925
 Name & Badge Number (Print)


 Signature

4/10/14
 Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Marlboro Twp. P.D., Marlboro, NJ

Date Received: 03/19/14

Date given to service: 03/19/14

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XA-0069

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFA-1895

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 04/03/14

Packing Slip

Customer no Order No Order date
 150062748 10691890 04/03/2014

Packing slip no. Ship date
 80848943 04/03/2014

Please reference on inquiries

ship to

150062749
 MARLBORO TOWNSHIP POLICE DEPARTMENT
 ATTN: ROSE MARIE DI LIBERTO
 MONMOUTH COUNTY
 1979 TOWNSHIP DRIVE
 MARLBORO, NJ 07746
 USA

bill to

Your Purchase Order Number

SGT. BRYAN OHARE
 Any questions? Please contact:
 CHARLES ROBINSON II

Your sales person

Don Pouliot
 Phone: 973-398-3228

Customer

MARBORO TOWNSHIP POLICE
 DEPARTMENT
 ATTN: ROSE MARIE DI LIBERTO
 MONMOUTH COUNTY
 1979 TOWNSHIP DRIVE
 MARLBORO, NJ 07746
 USA

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARXA-0069 RECEIVED 03/19/2014			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150062748 10691890 04/03/2014

Packing slip no. Ship date
80848943 04/03/2014

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	2 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Marlboro Twp. Police Dept. 1979 Township Drive Marlboro, NJ 07746		2. Contact: Lt. Joe Tomazic	4. Date: 5/11/2011
		3. Phone Number: (732) 536-0100	5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXA-0069</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Returned from Draeger

10. Comments:
Instrument was put back into service and a Calibration was performed on same with positive results.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Marc Dennis #5925

 Name & Badge Number (Print)



 Signature

5/11/11

 Date

Packing Slip

Customer no Order No Order date
 150062748 10473992 05/04/2011

Packing slip no. Ship date
 80597946 05/04/2011

Please reference on inquiries

ship to

150062749
 MARLBORO TOWNSHIP POLICE DEPARTMENT
 ATTN: DEPUTY CHEIF ERIC LANDAU
 MONMOUTH COUNTY
 1979 TOWNSHIP DRIVE
 MARLBORO, NJ 07746
 USA

bill to

Your Purchase Order Number

7110 WARRANTY: ARXA-0069
 Any questions? Please contact:
 MARA HERRERA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Customer

MARBORO TOWNSHIP POLICE
 DEPARTMENT
 ATTN: CAPTAIN STEVEN MENNONA
 MONMOUTH COUNTY
 1979 TOWNSHIP DRIVE
 MARLBORO, NJ 07746
 USA

Delivery terms

FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARXA-0069			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED			
0020	1 EA	6808486 PLATES, 7110 Ordered / Back ordered : 0 / 0 EA REPLACED			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150062748 10473992 05/04/2011

Packing slip no. Ship date
80597946 05/04/2011

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Marlboro Twp. Police Dept. 1979 Township Drive Marlboro, NJ 07746	2. Contact: Lt. Joe Tomazic	4. Date: 4/27/2011
	3. Phone Number: (732) 536-0100	5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXA-0069</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
Complaint received of instrument experiencing repeated control test failures.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several repeated control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Marc Dennis #5925
Tpr. I [Signature]
4/27/11

Name & Badge Number (Print)
Signature
Date

TO: _____ FAX: _____

Please use this form every time you ship Alcotest components to insure that they are serviced promptly and returned to your office

DRAEGER SAFETY EQUIPMENT RETURN FORM

DATE: 10/28/2008 CONTACT NAME Deputy Chief Landau

CUSTOMER BILLING NAME AND ADDRESS PO # Under Warranty

Under Warranty

CUSTOMER SHIP TO NAME AND ADDRESS

Marlboro Township Police Department
Att: Deputy Chief Eric Landau
1979 Township Drive
Marlboro, NJ 07746

PHONE (732)536-0100

FAX (732)972-7522

REASON FOR SERVICE RETURN: DST Clock _____

ANNUAL CALIBRATION / CERTIFICATION REPAIR X

PLEASE GIVE DESCRIPTION OF ANY PROBLEM(S) YOU ARE HAVING WITH YOUR EQUIPMENT.

032 Error Message.

SERIAL NUMBERS BEING SENT IN

ARXA-0069

PLEASE INSURE YOUR EQUIPMENT, AND SEND TO US BY UPS OR FED EX. WE WILL RETURN YOUR EQUIPMENT BY THE SAME SHIPPING METHOD AS IT WAS SHIPPED TO US. PLEASE INCLUDE THIS FORM IN PACKAGE WITH EQUIPMENT. REMEMBER: NEVER SEND IN YOUR DRY GAS WITH YOUR EQUIPMENT.

SHIP TO:

DRAEGER SAFETY DIAGNOSTICS, INC
4040 WEST ROYAL LANE SUITE # 136
IRVING, TX 75063

PHONE: 866-385-5900
FAX 972-929-1260

Packing Slip

Customer no Order No Order date
 150062748 10312741 11/05/2008

Packing slip no. **Ship date**
 80404559 11/05/2008

Please reference on inquiries

ship to
 150062749
 MARLBORO TOWNSHIP POLICE DEPARTMENT
 ATTN: LT. BARBELLA
 1979 TOWNSHIP DRIVE
 MONMOUTH COUNTY
 MARLBORO, NJ 07746
 USA

bill to

Your Purchase Order Number

LT. BARBELLA
 Any questions? Please contact:
 ERIKA ESQUEDA

Customer
 MARLBORO TOWNSHIP POLICE
 DEPARTMENT
 ATTN: DEPUTY CHEIF ERIC LANDAU
 1979 TOWNSHIP DRIVE
 MONMOUTH COUNTY
 MARLBORO, NJ 07746
 USA

Your sales person
 Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: FEDEX STD OVERNITE			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110 SERIAL #: ARXA-0069 ERROR 32- MOTOR FAULTY- REPLACED/ FC IS WJ SERIES- REPLACE AS PM- CAL W/ QC & OPS CHECKS			
0010	1 EA	6808374 GEARED ENGINE Ordered / Back ordered : 0 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150062748 10312741 11/05/2008

Packing slip no. Ship date
80404559 11/05/2008

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			
0040	1 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0 / 0 H			

Return & Repair Form

Customer Information:

Company Name: MARLBORO T/S B: _____ S: _____
 Date Received: 11-3-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - (NDA) - 2DY - 3DY - Other

Product Information:

Product: 7410 - (7110) - 6510 - 6810 Serial #: AR YA - 0069
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC? - ERROR 32 NOTED

Part Number	Description	Qty	Total Cost
<u>6808374</u>	<u>GEARED ENGINE</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>M.P. CAC 71</u>	<u>CAC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPL LABOR</u>		<u>1 HR</u>	<u>N/C WARR</u>

Repair Notes:

ERROR 32 - MOTOR FAULTY - REPAIRED / FC IS WJ SERIES -
REPLACE AS P.M. - CAC W/ QC OPS CHECKS

Service Technician: [Signature]

Date: 10-04-08

Return & Repair Form:

Customer Information:

Company Name: Marlboro Trup B: _____ S: _____
 Date Received: 5-2-11 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Serial #: AR XA - 0069
 Printer Ser #: AR
 Whole Instrument _____ Top 1/2 _____
 Sim Ser #: _____
 Other _____ Probe Ser # DD P
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>MP CGL 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1.5</u>	<u>N/C W</u>

Repair Notes:

FC# ARCA-0141
Cal w/ QC & Ops check

Service Technician: JS

Date: 5-4-2011