

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00512	Calib. Date:	10/09/2014
Certification File No.:	00513	Cert. Date:	10/09/2014
Linearity File No.:	00515	Lin. Date:	10/09/2014
Solution File No.:	00533	Soln. Date:	02/02/2015
Sequential File No.:	00533	File Date:	02/02/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1238

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:16S		
Control 1 EC	0.100%	13:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18S		
Control 2 EC	0.100%	13:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:19S		
Control 3 EC	0.100%	13:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-080 (14)

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SGT. R. J. Dennis #5925

Badge No.: 5925

Date: 02/02/2015