

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00041
Calibration File No.:	00788	Cert. Date:	04/15/2019
Certification File No.:	00789	Lin. Date:	04/15/2019
Linearity File No.:	00790	Soln. Date:	05/09/2019
Solution File No.:	00792	File Date:	05/09/2019
Sequential File No.:	00792		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0201

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:44D		
Control 1 EC	0.098%	09:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45D		
Control 2 EC	0.098%	09:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		
Control 3 EC	0.098%	09:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: _____



90

Badge No.: 090

Date: 05/09/2019

