Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE				
Calibration File No.:	00633	Calib. Date:	09/22/2016	Calib. No.:	00033
Certification File No.:	00634	Cert. Date:	09/22/2016	Cert. No.:	00024
Linearity File No.:	00635	Lin. Date:	09/22/2016	Lin. No.:	00028
Solution File No.:	00645	Soln. Date:	11/26/2016	Soln. No.:	00148
Sequential File No.:	00645	File Date:	11/26/2016		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0162
Control Solution %:	0.100%			Expires:	04/13/2017
Solution Control Lot:	15D138			Bottle No.:	0352

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:31S		
Control 1 EC	0.099%	12:32S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:32S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32S		
Control 2 EC	0.098%	12:33S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:33S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34S		
Control 3 EC	0.098%	12:34S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:34S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TISCHLER First Name: MICHAEL MI: J

Signature: Sgt. M. Tischler Badge No.: 075

Date: 11/26/2016

NOV 30 2016