

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00767 Calib. Date: 12/06/2018 Calib. No.: 00039  
Certification File No.: 00768 Cert. Date: 12/06/2018 Cert. No.: 00030  
Linearity File No.: 00769 Lin. Date: 12/06/2018 Lin. No.: 00034  
Solution File No.: 00771 Soln. Date: 12/10/2018 Soln. No.: 00191  
Sequential File No.: 00771 File Date: 12/10/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 09/14/2019  
Solution Control Lot: 17290 Bottle No.: 0636

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:05S		
Control 1 EC	0.099%	10:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		
Control 2 EC	0.097%	10:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08S		
Control 3 EC	0.097%	10:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: \_\_\_\_\_

#90

Badge No.: 090

Date: 12/10/2018

