

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00788 Calib. Date: 04/15/2019 Calib. No.: 00041
Certification File No.: 00789 Cert. Date: 04/15/2019 Cert. No.: 00031
Linearity File No.: 00790 Lin. Date: 04/15/2019 Lin. No.: 00035
Solution File No.: 00805 Soln. Date: 07/10/2019 Soln. No.: 00203
Sequential File No.: 00805 File Date: 07/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0040

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:03D		
Control 1 EC	0.096%	15:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	15:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:04D		
Control 2 EC	0.096%	15:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	15:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06D		
Control 3 EC	0.097%	15:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TISCHLER

First Name: MICHAEL

MI: J

Signature: Sgt. M. J. 075

Badge No.: 075

Date: 07/10/2019

