

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00788 Calib. Date: 04/15/2019 Calib. No.: 00041
Certification File No.: 00789 Cert. Date: 04/15/2019 Cert. No.: 00031
Linearity File No.: 00790 Lin. Date: 04/15/2019 Lin. No.: 00035
Solution File No.: 00813 Soln. Date: 08/26/2019 Soln. No.: 00209
Sequential File No.: 00813 File Date: 08/26/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0107

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:57D		
Control 1 EC	0.095%	10:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		
Control 2 EC	0.096%	10:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	10:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00D		
Control 3 EC	0.096%	11:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	11:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: *CM M*

#69

Badge No.: 069

Date: 08/26/2019

