Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE				
Calibration File No.:	00413	Calib. Date:	04/04/2013	Calib. No.:	00019
Certification File No.:	00414	Cert. Date:	04/04/2013	Cert. No.:	00012
Linearity File No.:	00415	Lin. Date:	04/04/2013	Lin. No.:	00015
Solution File No.:	00436	Soln. Date:	08/14/2013	Soln. No.:	00095
Sequential File No.:	00436	File Date:	08/14/2013		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0162
Control Solution %:	0.100%			Expires:	06/28/2014
Solution Control Lot:	12F102			Bottle No.:	0389

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:08D		
Control 1 EC	0.097%	11:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09D		
Control 2 EC	0.096%	11:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11D		
Control 3 EC	0.096%	11:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE First Name: BRYAN MI: F

Signature: #096 Badge No.: 96

Date: 08/14/2013

RECEIVED AUG 14 2013 MARLBORO TOWNSHIP POLICE DEPARTMENT