

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00481	Calib. Date:	04/10/2014
Certification File No.:	00482	Cert. No.:	00023
Linearity File No.:	00483	Cert. Date:	04/10/2014
Solution File No.:	00495	Lin. No.:	00018
Sequential File No.:	00495	Soln. Date:	06/01/2014
		Soln. No.:	00111
		File Date:	06/01/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1236

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:51D		
Control 1 EC	0.096%	06:52D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:52D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:52D		
Control 2 EC	0.097%	06:53D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	06:53D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:54D		
Control 3 EC	0.097%	06:54D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.096%	06:54D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WILLETT

First Name: BRIAN

MI: W

Signature:  #110

Badge No.: 110

Date: 06/01/2014

