

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00633 Calib. Date: 09/22/2016 Calib. No.: 00033
Certification File No.: 00634 Cert. Date: 09/22/2016 Cert. No.: 00024
Linearity File No.: 00635 Lin. Date: 09/22/2016 Lin. No.: 00028
Solution File No.: 00640 Soln. Date: 11/05/2016 Soln. No.: 00147
Sequential File No.: 00640 File Date: 11/05/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 04/13/2017
Solution Control Lot: 15D138 Bottle No.: 3580

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:38S		
Control 1 EC	0.098%	21:39S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:39S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40S		
Control 2 EC	0.098%	21:40S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:40S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:41S		
Control 3 EC	0.097%	21:42S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:42S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BISE

First Name: JOHN

MI: E

Signature: Sgt. John Bise #60

Badge No.: 60

Date: 11/05/2016

NOV 07 2016

NOV 07 2016