Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE				
Calibration File No.:	00633	Calib. Date:	09/22/2016	Calib. No.:	00033
Certification File No.:	00634	Cert. Date:	09/22/2016	Cert. No.:	00024
Linearity File No.:	00635	Lin. Date:	09/22/2016	Lin. No.:	00028
Solution File No.:	00636	Soln. Date:	09/22/2016	Soln. No.:	00145
Sequential File No.:	00636	File Date:	09/22/2016		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0162
Control Solution %:	0.100%			Expires:	04/13/2017
Solution Control Lot:	15D138			Bottle No.:	0365

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:29D		
Control 1 EC	0.099%	10:29D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:29D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		
Control 2 EC	0.098%	10:30D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:30D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31D		
Control 3 EC	0.098%	10:32D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:32D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDU5P2-066 (DK)

Changed By:

Last Name: KLIMIK First Name: DAVID MI: W

Signature: IPR D. K #7040 Badge No.: 7040

Date: 09/22/2016