

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00849 Calib. Date: 02/11/2020 Calib. No.: 00043
Certification File No.: 00850 Cert. Date: 02/11/2020 Cert. No.: 00033
Linearity File No.: 00851 Lin. Date: 02/11/2020 Lin. No.: 00037
Solution File No.: 00856 Soln. Date: 02/29/2020 Soln. No.: 00220
Sequential File No.: 00856 File Date: 02/29/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 06/10/2021
Solution Control Lot: 19150 Bottle No.: 0931

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:04S		
Control 1 EC	0.097%	06:05S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:05S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:06S		
Control 2 EC	0.098%	06:06S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	06:06S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:07S		
Control 3 EC	0.097%	06:08S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	06:08S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:08S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ARONNE

First Name: GREGORY

MI: A

Signature: 

Badge No.: 106

Date: 02/29/2020