

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00965	Calib. Date: 12/20/2021
Certification File No.:	00966	Calib. No.: 00048
Linearity File No.:	00967	Cert. Date: 12/20/2021
Solution File No.:	00985	Cert. No.: 00038
Sequential File No.:	00985	Lin. Date: 12/20/2021
		Lin. No.: 00042
		Soln. Date: 03/16/2022
		Soln. No.: 00264
		File Date: 03/16/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1369

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:17S		
Control 1 EC	0.100%	15:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19S		
Control 2 EC	0.100%	15:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:20S		
Control 3 EC	0.100%	15:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETERSON

First Name: BRIAN

MI: D

Signature: \_\_\_\_\_

Badge No.: 129

Date: 03/16/2022

