

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00965 Calib. Date: 12/20/2021 Calib. No.: 00048  
Certification File No.: 00966 Cert. Date: 12/20/2021 Cert. No.: 00038  
Linearity File No.: 00967 Lin. Date: 12/20/2021 Lin. No.: 00042  
Solution File No.: 00989 Soln. Date: 04/05/2022 Soln. No.: 00265  
Sequential File No.: 00989 File Date: 04/05/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 1341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:39D		
Control 1 EC	0.100%	16:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:40D		
Control 2 EC	0.101%	16:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:41D		
Control 3 EC	0.100%	16:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: PETERSON

First Name: BRIAN

MI: D

Signature: \_\_\_\_\_

*Brian Peterson* #129

Badge No.: 129

Date: 04/05/2022

