

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00927	Calib. Date: 03/09/2021
Certification File No.:	00928	Calib. No.: 00046
Linearity File No.:	00929	Cert. Date: 03/09/2021
Solution File No.:	00937	Cert. No.: 00036
Sequential File No.:	00937	Lin. Date: 03/09/2021
		Lin. No.: 00040
		Soln. Date: 04/14/2021
		Soln. No.: 00249
		File Date: 04/14/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:42D		
Control 1 EC	0.098%	15:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44D		
Control 2 EC	0.098%	15:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45D		
Control 3 EC	0.099%	15:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MORGANTE

First Name: MICHAEL

MI: A

Signature: _____



Badge No.: 128

Date: 04/14/2021

