

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00965 Calib. Date: 12/20/2021 Calib. No.: 00048
Certification File No.: 00966 Cert. Date: 12/20/2021 Cert. No.: 00038
Linearity File No.: 00967 Lin. Date: 12/20/2021 Lin. No.: 00042
Solution File No.: 00993 Soln. Date: 04/28/2022 Soln. No.: 00266
Sequential File No.: 00993 File Date: 04/28/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0623

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:45D		
Control 1 EC	0.101%	00:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:46D		
Control 2 EC	0.101%	00:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:47D		
Control 3 EC	0.101%	00:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	00:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: PETERSON

First Name: BRIAN

MI: D

Signature: _____



Badge No.: 129

Date: 04/28/2022

