

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARXA-0069 |
| Location: | MARLBORO TWP. POLICE | | |
| Calibration File No.: | 00864 | Calib. Date: | 05/04/2020 |
| Certification File No.: | 00865 | Cert. No.: | 00034 |
| Linearity File No.: | 00866 | Lin. Date: | 05/04/2020 |
| Solution File No.: | 00867 | Soln. No.: | 00225 |
| Sequential File No.: | 00867 | File Date: | 05/04/2020 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDXC S3-0164 |
| Solution Control Lot: | 19150 | Expires: | 06/10/2021 |
| | | Bottle No.: | 0936 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 11:09D | | |
| Control 1 EC | 0.100% | 11:10D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 11:10D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:11D | | |
| Control 2 EC | 0.099% | 11:11D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 11:11D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:12D | | |
| Control 3 EC | 0.100% | 11:13D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 11:13D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:13D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: _____

DDUJ2-080 (PC)

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: _____

TIC I Dent 7045

Badge No.: 7045

Date: 05/04/2020