

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00706 Calib. Date: 01/09/2018 Calib. No.: 00037
Certification File No.: 00707 Cert. Date: 01/09/2018 Cert. No.: 00028
Linearity File No.: 00708 Lin. Date: 01/09/2018 Lin. No.: 00032
Solution File No.: 00730 Soln. Date: 05/20/2018 Soln. No.: 00180
Sequential File No.: 00730 File Date: 05/20/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 0664

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:43D		
Control 1 EC	0.099%	06:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:44D		
Control 2 EC	0.098%	06:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:46D		
Control 3 EC	0.098%	06:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ARONNE

First Name: GREGORY

MI: A

Signature: _____



Badge No.: 106

Date: 05/20/2018

