

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00046
Calibration File No.:	00927	Cert. Date:	03/09/2021
Certification File No.:	00928	Lin. Date:	03/09/2021
Linearity File No.:	00929	Soln. Date:	05/31/2021
Solution File No.:	00948	File Date:	05/31/2021
Sequential File No.:	00948		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	1374

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:32D		
Control 1 EC	0.096%	11:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		
Control 2 EC	0.097%	11:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 3 EC	0.097%	11:35D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STACEY

First Name: BRANDON

MI: J

Signature: _____

Signature: SGT [Handwritten Signature] #78

Badge No.: 78

Date: 05/31/2021

