

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00927	Calib. Date:	03/09/2021
Certification File No.:	00928	Cert. Date:	03/09/2021
Linearity File No.:	00929	Lin. Date:	03/09/2021
Solution File No.:	00950	Soln. Date:	06/19/2021
Sequential File No.:	00950	File Date:	06/19/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	1373

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:37D		
Control 1 EC	0.101%	16:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38D		
Control 2 EC	0.099%	16:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:40D		
Control 3 EC	0.099%	16:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORGANTE

First Name: MICHAEL

MI: A

Signature: _____



Badge No.: 128

Date: 06/19/2021

