

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00864 Calib. Date: 05/04/2020 Calib. No.: 00044  
Certification File No.: 00865 Cert. Date: 05/04/2020 Cert. No.: 00034  
Linearity File No.: 00866 Lin. Date: 05/04/2020 Lin. No.: 00038  
Solution File No.: 00873 Soln. Date: 06/27/2020 Soln. No.: 00228  
Sequential File No.: 00873 File Date: 06/27/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164  
Control Solution %: 0.100% Expires: 06/10/2021  
Solution Control Lot: 19150 Bottle No.: 0934

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:16D		
Control 1 EC	0.097%	11:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		
Control 2 EC	0.097%	11:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		
Control 3 EC	0.097%	11:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: \_\_\_\_\_

Badge No.: 090

Date: 06/27/2020

