

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00890	Calib. Date:	09/18/2020
Certification File No.:	00891	Cert. Date:	09/18/2020
Linearity File No.:	00892	Lin. Date:	09/18/2020
Solution File No.:	00904	Soln. Date:	11/10/2020
Sequential File No.:	00904	File Date:	11/10/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	1390

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:56S		
Control 1 EC	0.100%	18:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:57S		
Control 2 EC	0.100%	18:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:58S		
Control 3 EC	0.100%	18:59S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: PETERSON

First Name: BRIAN

MI: D

Signature: \_\_\_\_\_

Badge No.: 129

Date: 11/10/2020

