

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
 Location: MARLBORO TWP. POLICE
 Calibration File No.: 00543 Calib. Date: 03/23/2015 Calib. No.: 00025
 Certification File No.: 00544 Cert. Date: 03/23/2015 Cert. No.: 00017
 Linearity File No.: 00545 Lin. Date: 03/23/2015 Lin. No.: 00021
 Solution File No.: 00546 Soln. Date: 03/23/2015 Soln. No.: 00122
 Sequential File No.: 00546 File Date: 03/23/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
 Control Solution %: 0.100% Expires: 11/12/2015
 Solution Control Lot: 13K125 Bottle No.: 1229

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:12S	17.12 D ^(1b)	
Control 1 EC	0.102%	16:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:14S		
Control 2 EC	0.102%	16:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15S		
Control 3 EC	0.102%	16:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17S	17.17 D ^(1b)	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDVJ P2-080 ^(1b)

Changed By:

Last Name: DENNIS First Name: MARC MI: W.

Signature: Sgt. [Signature] #5925 Badge No.: 5925
 Date: 03/23/2015

MAR 24 2015