

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calibration File No.:	00481
		Calib. Date:	04/10/2014
Calibration File No.:	00481	Calib. No.:	00023
Certification File No.:	00482	Cert. Date:	04/10/2014
Linearity File No.:	00483	Cert. No.:	00015
Solution File No.:	00506	Lin. Date:	04/10/2014
Sequential File No.:	00506	Lin. No.:	00018
		Soln. Date:	09/06/2014
		Soln. No.:	00114
		File Date:	09/06/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0162
Control Solution %:	0.100%			Expires:	11/12/2015
Solution Control Lot:	13K125			Bottle No.:	1237

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:43D		
Control 1 EC	0.097%	22:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:44D		
Control 2 EC	0.096%	22:45D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:45D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45D		
Control 3 EC	0.097%	22:46D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:46D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Lt. [Signature] #096

Badge No.: 96

Date: 09/06/2014

