

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Marlboro Twp. Police Department 1979 Township Drive Marlboro, NJ 07746</b>	2. Contact: <b>Sgt. Tim Snyder #105</b>	4. Date: <b>10/7/2015</b>
	3. Phone Number: <b>(732) 536-0100 Ext. 1105</b>	5. County: <b>Monmouth</b>

6. Alcotest Instrument Serial Number: <b>ARXA-0069</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>
---	---	---


9. Reason for Service:  
**Alcotest instrument returned from outside evaluation or repair and placed back into service.**

10. Comments:  
**Instrument was put back into service and a Calibration was performed on same with positive results. The instrument was returned with a packing slip, Draeger Return and Repair Form, and a New Certificate of Accuracy.**

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Sgt. Marc Dennis #5925**  
 Name & Badge Number (Print)
 

  
 Signature
 

**10/7/15**  
 Date

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Marlboro Twp. P.D., Marlboro, NJ

Date Received: 09/16/15

Date given to service: 09/16/15

Carrier:  FedEx  UPS  USPS      Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810      Serial Number: AR XA-0069  
 7110  8610      Printer Serial#: AR  
 7510  DT5000      Sim Serial#: DD  
 7410 Upper-half  9510      Probe Serial#: DD  
 7410 Whole

Warranty Expires: November 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter       Regulator       Printer Ribbon       Printer Paper  
 Mouthpieces       9510 Stylus       9510 Top Cover       Carrying Case  
 Dry Gas       Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W

Repair Notes: Replaced defective fuel cell.

\_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Fuel Cell = ARHE-1986

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 09/21/15

# Packing Slip

Customer no      Order No      Order date  
 150062748    10804695    09/21/2015

Packing slip no.      Ship date  
 80977619      09/21/2015

Please reference on inquiries

Ship to  
 150062749  
 MARLBORO TOWNSHIP POLICE DEPARTMENT  
 ATTN: ROSE MARIE DI LIBERTO  
 MONMOUTH COUNTY  
 1979 TOWNSHIP DRIVE  
 MARLBORO, NJ 07746-2299  
 USA

Payer

Your Purchase Order Number

SGT. TIM SNYDER #105  
 Any questions? Please contact:  
 ROSEMARIE HUGHES

Customer  
 MARLBORO TOWNSHIP POLICE  
 DEPARTMENT  
 ATTN: ROSE MARIE DI LIBERTO  
 MONMOUTH COUNTY  
 1979 TOWNSHIP DRIVE  
 MARLBORO, NJ 07746-2299  
 USA

Your sales person

LINDA SALO  
 Phone: 866-385-5900  
 Fax: 972-929-1260

Delivery terms  
 FA  
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARXA-0069 REC'D 9/16/15			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 EA			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 536410  
 Pittsburgh, PA 15253-5906

# Packing Slip

Customer no    Order No    Order date  
150062748    10804695    09/21/2015

Packing slip no.    Ship date  
80977619    09/21/2015

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Marlboro Twp. Police Department 1979 Township Drive Marlboro, NJ 07746</b>		2. Contact: <b>Sgt. Tim Snyder #105</b>	4. Date: <b>9/10/2015</b>
		3. Phone Number: <b>(732) 536-0100 Ext. 1105</b>	5. County: <b>Monmouth</b>
6. Alcotest Instrument Serial Number: <b>ARXA-0069</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Was at agency to perform semi-annual calibration and during the Linearity Test it failed on test 5 during the 0.160 of the test which was displayed on the LED and printed the same on the printout indicating a low EC reading resulting in a failure.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument and due to the failed linearity test, the Instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

**Last Sequential File # 00574**

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<u>Sgt. Marc Dennis #5925</u> Name & Badge Number (Print)	<u>Sgt. M. Dennis</u> Signature	<u>9/10/15</u> Date
--	------------------------------------	------------------------

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Marlboro Twp. Police Dept.  1979 Township Drive  Marlboro, NJ 07746</b>		2. Contact: <b>Sgt. Bryan Ohare</b>		4. Date: <b>4/10/2014</b>
		3. Phone Number: <b>(732) 536-0100</b>		5. County: <b>Monmouth</b>
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARXA-0069</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>		


9. Reason for Service:  
**Returned from Draeger.**

10. Comments:  
**Instrument was put back into service and a Calibration was performed on same with positive results.**

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Sgt. Marc Dennis #5925**  
\_\_\_\_\_  
Name & Badge Number (Print)

  
\_\_\_\_\_  
Signature

**4/10/14**  
\_\_\_\_\_  
Date

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Marlboro Twp. P.D., Marlboro, NJ

Date Received: 03/19/14 Date given to service: 03/19/14

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR XA-0069  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Fuel Cell = ARFA-1895  
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 04/03/14

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Marlboro Twp. Police Dept. 1979 Township Drive Marlboro, NJ 07746</b>	2. Contact: <b>Sgt. Bryan Ohare</b>	4. Date: <b>3/13/2014</b>
	3. Phone Number: <b>(732) 536-0100</b>	5. County: <b>Monmouth</b>
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARXA-0069</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>

9. Reason for Service:  
**At above location to perform a calibration and noticed the EC numbers on the Part II Linearity Test were low and test 6 actually failed due to same and believed the Fuel Cell is depleting.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed a few repeated control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Sgt. Marc Dennis #5925**  
 Name & Badge Number (Print)

**SGT. [Signature]**  
 Signature

**3/13/14**  
 Date



# Packing Slip

Customer no    Order No    Order date  
 150062748    10691890    04/03/2014

Packing slip no.    Ship date  
 80848943            04/03/2014  
 Please reference on inquiries

ship to

150062749  
 MARLBORO TOWNSHIP POLICE DEPARTMENT  
 ATTN: ROSE MARIE DI LIBERTO  
 MONMOUTH COUNTY  
 1979 TOWNSHIP DRIVE  
 MARLBORO, NJ 07746  
 USA

bill to

Your Purchase Order Number

SGT. BRYAN OHARE  
 Any questions? Please contact:  
 CHARLES ROBINSON II

Customer  
 MARLBORO TOWNSHIP POLICE  
 DEPARTMENT  
 ATTN: ROSE MARIE DI LIBERTO  
 MONMOUTH COUNTY  
 1979 TOWNSHIP DRIVE  
 MARLBORO, NJ 07746  
 USA

Your sales person

Don Pouliot  
 Phone: 973-398-3228

Delivery terms  
 FA  
 FREIGHT ALLOWED

Page    1   /   2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARXA-0069 RECEIVED 03/19/2014			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150062748    10691890    04/03/2014

Packing slip no.    Ship date  
80848943    04/03/2014

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA
0040	2 EA	6808486 PLATES, 7110--9510 Ordered / Back ordered : 0 / 0 EA

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Marlboro Twp. Police Dept. 1979 Township Drive Marlboro, NJ 07746</b>	2. Contact: <b>Lt. Joe Tomazic</b>  3. Phone Number: <b>(732) 536-0100</b>	4. Date: <b>5/11/2011</b>  5. County: <b>Monmouth</b>
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARXA-0069</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>


9. Reason for Service:  
**Returned from Draeger**

10. Comments:  
**Instrument was put back into service and a Calibration was performed on same with positive results.**

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Tpr. I Marc Dennis #5925**  
 \_\_\_\_\_  
 Name & Badge Number (Print)

  
 \_\_\_\_\_  
 Signature

**5/11/11**  
 \_\_\_\_\_  
 Date

# Packing Slip


Customer no    Order No    Order date  
 150062748    10473992    05/04/2011

Packing slip no.    Ship date  
 80597946    05/04/2011  
 Please reference on inquiries

ship to

150062749  
 MARLBORO TOWNSHIP POLICE DEPARTMENT  
 ATTN: DEPUTY CHEIF ERIC LANDAU  
 MONMOUTH COUNTY  
 1979 TOWNSHIP DRIVE  
 MARLBORO, NJ 07746  
 USA

bill to



Your Purchase Order Number

7110 WARRANTY: ARXA-0069  
 Any questions? Please contact:  
 MARA HERRERA

Customer

MARLBORO TOWNSHIP POLICE  
 DEPARTMENT  
 ATTN: CAPTAIN STEVEN MENNONA  
 MONMOUTH COUNTY  
 1979 TOWNSHIP DRIVE  
 MARLBORO, NJ 07746  
 USA

Your sales person

Don Pouliot  
 Phone: 973-398-3228

Delivery terms

FA  
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARXA-0069			
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA REPLACED			
0020	1 EA	6808486 PLATES,7110 Ordered / Back ordered : 0 / 0 EA REPLACED			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337

# Packing Slip

Customer no    Order No    Order date  
150062748    10473992    05/04/2011

Packing slip no.    Ship date  
80597946    05/04/2011

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Marlboro Twp. Police Dept. 1979 Township Drive Marlboro, NJ 07746</b>		2. Contact: <b>Lt. Joe Tomazic</b>	4. Date: <b>4/27/2011</b>
		3. Phone Number: <b>(732) 536-0100</b>	5. County: <b>Monmouth</b>
6. Alcotest Instrument Serial Number: <b>ARXA-0069</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Complaint received of instrument experiencing repeated control test failures.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several repeated control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Tpr. I Marc Dennis #5925**  
Name & Badge Number (Print)

*Tpr. I [Signature]*      **4/27/11**  
Signature      Date

TO: \_\_\_\_\_ FAX: \_\_\_\_\_

**Please use this form every time you ship Alcotest components to insure that they are serviced promptly and returned to your office**

### DRAEGER SAFETY EQUIPMENT RETURN FORM

DATE: 10/28/2008 CONTACT NAME Deputy Chief Landau

CUSTOMER BILLING NAME AND ADDRESS PO # Under Warranty

Under Warranty

**CUSTOMER SHIP TO NAME AND ADDRESS**

Marlboro Township Police Department  
Att: Deputy Chief Eric Landau  
1979 Township Drive  
Marlboro, NJ 07746

PHONE (732)536-0100

FAX (732)972-7522

REASON FOR SERVICE RETURN: DST Clock \_\_\_\_\_

ANNUAL CALIBRATION / CERTIFICATION REPAIR X

PLEASE GIVE DESCRIPTION OF ANY PROBLEM(S) YOU ARE HAVING WITH YOUR EQUIPMENT.

032 Error Message.

SERIAL NUMBERS BEING SENT IN

ARXA-0069

**PLEASE INSURE YOUR EQUIPMENT, AND SEND TO US BY UPS OR FED EX. WE WILL RETURN YOUR EQUIPMENT BY THE SAME SHIPPING METHOD AS IT WAS SHIPPED TO US. PLEASE INCLUDE THIS FORM IN PACKAGE WITH EQUIPMENT. REMEMBER: NEVER SEND IN YOUR DRY GAS WITH YOUR EQUIPMENT.**

SHIP TO:

DRAEGER SAFETY DIAGNOSTICS, INC  
4040 WEST ROYAL LANE SUITE # 136  
IRVING, TX 75063

PHONE: 866-385-5900  
FAX 972-929-1260

# Packing Slip

Customer no    Order No    Order date  
 150062748    10312741    11/05/2008

Packing slip no.    Ship date  
 80404559    11/05/2008

Please reference on inquiries

ship to  
 150062749  
 MARLBORO TOWNSHIP POLICE DEPARTMENT  
 ATTN: LT. BARBELLA  
 1979 TOWNSHIP DRIVE  
 MONMOUTH COUNTY  
 MARLBORO, NJ 07746  
 USA

bill to

Your Purchase Order Number

LT. BARBELLA  
 Any questions? Please contact:  
 ERIKA ESQUEDA

Customer  
 MARLBORO TOWNSHIP POLICE  
 DEPARTMENT  
 ATTN: DEPUTY CHEIF ERIC LANDAU  
 1979 TOWNSHIP DRIVE  
 MONMOUTH COUNTY  
 MARLBORO, NJ 07746  
 USA

Your sales person  
 Don Pouliot  
 Phone: 973-398-3228

Delivery terms  
 FA  
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: FEDEX STD OVERNITE
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110 SERIAL #: ARXA-0069 ERROR 32- MOTOR FAULTY- REPLACED/ FC IS WJ SERIES- REPLACE AS PM- CAL W/ QC & OPS CHECKS
0010	1 EA	6808374 GEARED ENGINE Ordered / Back ordered : 0 / 0 EA

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 200337  
 Pittsburgh, PA 15251-0337



# Packing Slip

Customer no    Order No    Order date  
150062748    10312741    11/05/2008

Packing slip no.    Ship date  
80404559    11/05/2008

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110 Ordered / Back ordered : 0 / 0 EA			
0040	1 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0 / 0 H			

# Return & Repair Form

## Customer Information:

Company Name: MARLBORO T/S B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 11-3-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - (NDA) - 2DY - 3DY - Other

## Product Information:

Product: 7410 - (7110) - 6510 - 6810 Serial #: AR XA - 0069  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC? - ERROR 32 NOTED

Part Number	Description	Qty	Total Cost
<u>6808374</u>	<u>GEARED ENGINE</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>IMP CAC 71</u>	<u>CAC</u>	<u>1</u>	<u>N/C WARR</u>
<u>IMP LABOR</u>		<u>1 HR</u>	<u>N/C WARR</u>

### Repair Notes:

ERROR 32 - MOTOR FAULTY - REPAIRED / FC IS WJ SERIES -  
REPLACE AS PM - CAC W/ QC OPS CHECKS

Service Technician: [Signature]

Date: 10-04-08

# Return & Repair Form:

## Customer Information:

Company Name: Marlboro Twp B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 5-2-11 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: AR XA - 0069  
 Printer Ser #: AR  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Reason for Return:		Test #	
<u>FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C w</u>
<u>MP CAL 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C w</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1.5</u>	<u>N/C w</u>

### Repair Notes:

FC# ARCA-0141  
Cal w/ QC & Ops check

Service Technician: JS

Date: 5-4-2011