

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00368	Calib. Date:	10/12/2012
Certification File No.:	00369	Cert. Date:	10/12/2012
Linearity File No.:	00371	Lin. Date:	10/12/2012
Solution File No.:	00396	Soln. Date:	02/08/2013
Sequential File No.:	00396	File Date:	02/08/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0379

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:15S		
Control 1 EC	0.099%	10:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17S		
Control 2 EC	0.098%	10:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		
Control 3 EC	0.098%	10:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_

*Sgt. Jim #016*

Badge No.: 96

Date: 02/08/2013

