

# Calibrating Unit

## New Standard Solution Report

|                         |                       |              |              |
|-------------------------|-----------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARXA-0069    |
| Location:               | MARLBORO TWP. POLICE  |              |              |
| Calibration File No.:   | 00368                 | Calib. Date: | 10/12/2012   |
| Certification File No.: | 00369                 | Cert. Date:  | 10/12/2012   |
| Linearity File No.:     | 00371                 | Lin. Date:   | 10/12/2012   |
| Solution File No.:      | 00396                 | Soln. Date:  | 02/08/2013   |
| Sequential File No.:    | 00396                 | File Date:   | 02/08/2013   |
|                         |                       |              |              |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                | Serial No.:  | DDXC S3-0162 |
| Solution Control Lot:   | 12F102                | Expires:     | 06/28/2014   |
|                         |                       | Bottle No.:  | 0379         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:15S |                |                     |
| Control 1 EC      | 0.099% | 10:16S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 10:16S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:17S |                |                     |
| Control 2 EC      | 0.098% | 10:17S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.098% | 10:17S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:18S |                |                     |
| Control 3 EC      | 0.098% | 10:19S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 10:19S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:20S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_

*Sgt. Jim #016*

Badge No.: 96

Date: 02/08/2013

