

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00019
Calibration File No.:	00413	Calib. Date:	04/04/2013
Certification File No.:	00414	Cert. Date:	04/04/2013
Linearity File No.:	00415	Lin. Date:	04/04/2013
Solution File No.:	00436	Soln. Date:	08/14/2013
Sequential File No.:	00436	File Date:	08/14/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0389

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:08D		
Control 1 EC	0.097%	11:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09D		
Control 2 EC	0.096%	11:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11D		
Control 3 EC	0.096%	11:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE	First Name: BRYAN	MI: F
Signature: <u></u>	Badge No.: 96	Date: 08/14/2013

