

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00368 Calib. Date: 10/12/2012 Calib. No.: 00018
Certification File No.: 00369 Cert. Date: 10/12/2012 Cert. No.: 00011
Linearity File No.: 00371 Lin. Date: 10/12/2012 Lin. No.: 00014
Solution File No.: 00389 Soln. Date: 01/12/2013 Soln. No.: 00084
Sequential File No.: 00389 File Date: 01/12/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0378

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:17S		
Control 1 EC	0.100%	10:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		
Control 2 EC	0.098%	10:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:20S		
Control 3 EC	0.099%	10:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Sgt. Bryan OHare #2016

Badge No.: 96

Date: 01/12/2013

