

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00368 Calib. Date: 10/12/2012 Calib. No.: 00018
Certification File No.: 00369 Cert. Date: 10/12/2012 Cert. No.: 00011
Linearity File No.: 00371 Lin. Date: 10/12/2012 Lin. No.: 00014
Solution File No.: 00411 Soln. Date: 03/30/2013 Soln. No.: 00088
Sequential File No.: 00411 File Date: 03/30/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0387

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:28S		
Control 1 EC	0.100%	05:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:30S		
Control 2 EC	0.098%	05:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:31S		
Control 3 EC	0.099%	05:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LAYMAN

First Name: JOSHUA

MI: E

Signature: _____



Badge No.: 101

Date: 03/30/2013

