

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00032
Calibration File No.:	00608	Calib. Date:	03/23/2016
Certification File No.:	00609	Cert. Date:	03/23/2016
Linearity File No.:	00610	Lin. Date:	03/23/2016
Solution File No.:	00619	Soln. Date:	06/07/2016
Sequential File No.:	00619	File Date:	06/07/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	0371

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:13D		
Control 1 EC	0.100%	12:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		
Control 2 EC	0.099%	12:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16D		
Control 3 EC	0.099%	12:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____



Badge No.: 96

Date: 06/07/2016

