

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00608	Calib. Date: 03/23/2016
Certification File No.:	00609	Calib. No.: 00032
Linearity File No.:	00610	Cert. Date: 03/23/2016
Solution File No.:	00623	Cert. No.: 00023
Sequential File No.:	00623	Lin. Date: 03/23/2016
		Lin. No.: 00027
		Soln. Date: 07/05/2016
		Soln. No.: 00142
		File Date: 07/05/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 0351

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:01D		
Control 1 EC	0.100%	22:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:03D		
Control 2 EC	0.099%	22:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:04D		
Control 3 EC	0.100%	22:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TISCHLER

First Name: MICHAEL

MI: J

Signature: Sgt. M. J. Tischler 025

Badge No.: 075

Date: 07/05/2016

