

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00413	Calib. Date: 04/04/2013
Certification File No.:	00414	Cert. No.: 00019
Linearity File No.:	00415	Cert. Date: 04/04/2013
Solution File No.:	00418	Lin. No.: 00012
Sequential File No.:	00418	Lin. Date: 04/04/2013
		Lin. No.: 00015
		Soln. Date: 04/30/2013
		Soln. No.: 00090
		File Date: 04/30/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0380

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:14D		
Control 1 EC	0.097%	10:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		
Control 2 EC	0.096%	10:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 3 EC	0.096%	10:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_

*Sgt. [Signature] #096*

Badge No.: 96

Date: 04/30/2013

